



(1) **EC-TYPE EXAMINATION CERTIFICATE**

(2) **Equipment and protective systems intended for use in potentially explosive atmospheres - Directive 94/9/EC**

(3) EC-Type Examination Certificate Number: **KEMA 09ATEX0044 X** Issue Number: **1**

(4) Equipment: **Electro Magnetic Flowmeter Primary Head Series VersaFlow Mag 2000 F**

(5) Manufacturer: **Honeywell International, HFS**

(6) Address: **512 Virginia Drive, Fort Washington, PA 19034, USA**

(7) This equipment and any acceptable variation thereto is specified in the schedule to this certificate and the documents therein referred to.

(8) KEMA Quality B.V., notified body number 0344 in accordance with Article 9 of the Council Directive 94/9/EC of 23 March 1994, certifies that this equipment has been found to comply with the Essential Health and Safety Requirements relating to the design and construction of equipment and protective systems intended for use in potentially explosive atmospheres given in Annex II to the directive.

The examination and test results are recorded in confidential test report number 212419200/1.

(9) Compliance with the Essential Health and Safety Requirements has been assured by compliance with:

<b>EN 50014 : 1997 + A1, A2</b>	<b>EN 50018 : 2000 +A1</b>	<b>EN 50019 : 2000</b>
<b>EN 50020 : 2002</b>	<b>EN 50028 : 1987</b>	<b>EN 50281-1-1 : 1998 + A1</b>
<b>EN 60079-0 : 2004</b>	<b>EN 60079-18 : 2004</b>	

(10) If the sign "X" is placed after the certificate number, it indicates that the equipment is subject to special conditions for safe use specified in the schedule to this certificate.

(11) This EC-Type Examination Certificate relates only to the design, examination and tests of the specified equipment according to the Directive 94/9/EC. Further requirements of the directive apply to the manufacturing process and supply of this equipment. These are not covered by this certificate.

(12) The marking of the equipment shall include the following:



<b>II 2 GD</b>	<b>EEx me ia IIC or EEx de ia IIC T6...T3</b>	<b>(series without flange)</b>
	<b>T85 °C ... T180 °C</b>	
<b>II 2 GD</b>	<b>EEx e ia mb IIC T6...T3</b>	<b>(series with flange)</b>
	<b>T85 °C ... T180 °C</b>	

This certificate is issued on June 23, 2009 and, as far as applicable, shall be revised before the date of cessation of presumption of conformity of (one of) the standards mentioned above as communicated in the Official Journal of the European Union.

KEMA Quality B.V.

H.J.G. de Wild  
Certification Manager



\*\* Integral publication of this certificate and adjoining reports is allowed. This Certificate may only be reproduced in its entirety and without any change.



(13) **SCHEDULE**

(14) **to EC-Type Examination Certificate KEMA 09ATEX0044 X** Issue No. 1

(15) **Description**

The Electro Magnetic Flowmeter Primary Head, series VersaFlow Mag 2000 F is used to convert the flow of a conducting fluid into an electrical signal. An associated flow transmitter is used to supply the field coils of the primary head and to convert the measured electrode signal into an output signal.

The degree of protection of the apparatus enclosure is at least IP64 according to EN 60529.

Electro Magnetic Flowmeter Primary Head Series without Flange

The type of protection for the various sizes of primary heads is as follows:  
 DN2.5 - DN15: encapsulation "m" (field coils) and increased safety "e" (field coil wiring) or  
 DN25 - DN100: flameproof enclosure "d" (field coils and field coil wiring).

The terminal compartment is in type of protection increased safety "e".

The electrodes are in type of protection intrinsic safety "i".

Ambient temperature range -20 °C to +65 °C (sizes DN2.5 - DN15).  
 Ambient temperature range -40 °C to +65 °C (sizes DN25 - DN100).  
 Process temperature range -20 °C to +180 °C (sizes DN2.5 - DN15).  
 Process temperature range -40 °C to +180 °C (sizes DN25 - DN100).

The relation between temperature class, maximum surface temperature, maximum process temperature and ambient temperature is shown in the following table:

Temperature class	Max. surface temperature	Max. process temperature		
		Ta ≤ 40 °C	40 °C < Ta ≤ 50 °C	50 °C < Ta ≤ 65 °C
T6	T85 °C	65 °C	65 °C	60 °C
T5	T95 °C	85 °C	85 °C	75 °C
T4	T130 °C	125 °C	125 °C	115 °C
T3	T180 °C	180 °C	165 °C	140 °C

The maximum surface temperature T is based on an ambient temperature of 65 °C.

Electro Magnetic Flowmeter Primary Head Series with Flange

The Electro Magnetic Flowmeter Primary Head sizes DN15, DN25, DN40, DN50, DN80 and DN100 are manufactured with stainless- or carbon steel flanges, welded to the coil housing and with ceramic tubing encased.

The type and level of protection of the flowtube for all sizes is encapsulation "mb".

The terminal compartment is in type of protection increased safety "e".

The electrodes are in type of protection intrinsic safety "i".

Ambient temperature range for all sizes of the VersaFlow with flanges: -40 °C to +65 °C.



(13) **SCHEDULE**

(14) **to EC-Type Examination Certificate KEMA 09ATEX0044 X** Issue No. 1

Electric strength tests according to EN 50019 Clause 6.1 shall be applied during one minute without breakdown as follows:

- On each terminal compartment with 500 V between the field coil circuit and the enclosure and with 1500 V between the field coil circuit and the intrinsically safe sensor circuit.
- On each primary head with field coil wiring in type of protection increased safety "e" with 500 V between the field coils circuit and the enclosure and with 1500 V between the field coils circuit and the intrinsically safe sensor circuit.

Routine tests according to EN 50028 must be carried out on the primary heads with field coils in type of protection encapsulation "m" as follows:

- Clause 7.1: Visual check;
- Clause 7.2: Each primary head shall withstand a test voltage of 1500 V during one minute without breakdown between the field coils circuit and the enclosure and between the field coils circuit and the intrinsically safe sensor circuit;
- Clause 7.3: Check of the electrical data.

Electro Magnetic Flowmeter Primary Head Series with Flange

Electric strength tests according to EN 50019 Clause 6.1 shall be applied during one minute without breakdown as follows:

- On each terminal compartment with 500 V between the field coil circuit and the enclosure and with 1500 V between the field coil circuit and the intrinsically safe sensor circuit.

Routine tests according to EN 60079-18 must be carried out on the primary heads with field coils in type of protection encapsulation "m" as follows:

- Clause 9.1: Visual check

(16) **Test Report**

KEMA No. 212419200/1.

(17) **Special conditions for safe use**

The relation between temperature class, maximum surface temperature, maximum process temperature and ambient temperature is as shown above in description (15).

The field coils in type of protection "m" must be protected by a 160 mA fuse. The breaking capacity of the fuse must be in accordance with the prospective short circuit current of the supply.

Electro Magnetic Flowmeter Primary Head Series without Flange

A heat resistant cable with a continuous operating temperature of at least 85 °C must be used when:

- $T_a \leq 40 \text{ °C}$  and process temperature  $\geq 165 \text{ °C}$ , or
- $40 \text{ °C} < T_a \leq 50 \text{ °C}$  and process temperature  $\geq 130 \text{ °C}$ , or
- $T_a > 50 \text{ °C}$  and process temperature  $\geq 100 \text{ °C}$ .

Electro Magnetic Flowmeter Primary Head Series with Flange

For ambient temperatures above 40 °C suitable heat resistant cable shall be used in accordance with the instructions of the manufacturer.



(13) **SCHEDULE**

(14) **to EC-Type Examination Certificate KEMA 09ATEX0044 X Issue No. 1**

The relation between temperature class, maximum surface temperature, maximum process temperature and ambient temperature is shown in the following table:

Temperature class	Max. surface temperature	Max. process temperature		
		Ta ≤ 40 °C	40 °C < Ta ≤ 50 °C	50 °C < Ta ≤ 65 °C
T6	T85 °C	80 °C	80 °C	75 °C
T5	T100 °C	95 °C	95 °C	95 °C
T4	T135 °C	130 °C	130 °C	130 °C
T3	T180 °C	180 °C	180 °C	145 °C

The maximum surface temperature T is based on an ambient temperature of 65 °C.

**Electrical data**

Field coil circuit ..... U ≤ 40 V (pulsed)  
 (terminals 7, 8 and 9) I ≤ 125 mA (fuse protected)

Electrodes circuit (terminals 1, 2, 3 and 4):  
 in type of protection intrinsic safety EEx ia IIC, only for connection to a certified intrinsically safe circuit, with the following maximum values:  
 U<sub>i</sub> = 20 V; I<sub>i</sub> = 175 mA; C<sub>i</sub> = 0 nF; L<sub>i</sub> = 0 mH

The aforementioned intrinsically safe circuits shall, from the safety point of view, be considered to be connected to ground.

**Installation instructions**

When used in a potentially explosive atmosphere, requiring the use of apparatus of equipment category 2G, certified cable entry devices shall be used that are suitable for the application and correctly installed.

When used in a potentially explosive atmosphere, requiring the use of apparatus of equipment category 2D, certified cable entry devices with a degree of ingress protection of at least IP6X according to EN 60529 shall be used that are suitable for the application and correctly installed.

Unused openings shall be closed with suitable certified closing elements.

With the use of conduit, a suitable certified sealing device such as a stopping box with setting compound shall be provided immediately at the entrance to the flameproof enclosure.

**Routine tests**

Electro Magnetic Flowmeter Primary Head Series without Flange

An overpressure test according to EN 50018 clause 16 must be carried out on each welded primary head in type of protection flameproof enclosure "d" at a test pressure of 22,5 bar during at least 10 seconds.



(13) **SCHEDULE**

(14) **to EC-Type Examination Certificate KEMA 09ATEX0044 X** Issue No. 1

(18) **Essential Health and Safety Requirements**

Covered by the standards listed at (9).

(19) **Test documentation**

As listed in Test Report No. 212419200/1.